



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Weighing/Load Receiving Element
Load Cell Electronic
Model: LS-WIM Axle Scale
 n_{max} : 2000
 e_{min} : 20 lb
Capacity: Up to 40 000 lb
Platform: 2.5 ft x 13 ft

Accuracy Class: III

Submitted By:

Intercomp Company
3839 County Road 116
Medina, MN 55340
Tel: 763-476-2531 x321
Fax: 763-476-2613
Contact: Matt Young
Email: matt@intercompcompany.com
Web site: www.intercompcompany.com

Standard Features and Options

Standard Features:

- Scale Capacities: Up to 40 000 lb
- Platform Area: 32.5 sq. ft, minimum
- Platform Deck Material: Steel
- Pit Installation

Load Cells Used:

Intercomp Company, Model 603053-AC 25K Strain Gauge, Double-ended Shear Beam Load Cells (NTEP CC Number 94-121), or other NTEP Certified, Compatible and Equivalent.

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

James Cassidy

Chairman, NCWM, Inc.

Kristin Macey

Committee Chair, National Type Evaluation Program Committee

Issued: March 27, 2018

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Intercomp Company
Weighing/Load Receiving Element / LS-WIM Axle Scale

Application: For use in general purpose weighing when interfaced with an NTEP certified and compatible indicating element.

Identification: The identification badge is located under the junction box access cover and is mounted to the junction box lid.

Sealing: The indicating element and junction box shall be sealed according to manufacturer specifications.

Test Conditions: This Certificate supersedes Certificate of Conformance 15-098 and was issued to change the Accuracy Class designation. NCWM Publication 14, Weighing Devices, Technical Policy Previous prohibits vehicle scale weighing / load receiving elements lengths shorter than 7 ft. NTEP test data was reviewed by NTEP to verify compliance with the accuracy class. Previous test conditions are listed below for reference.

Certificate of Conformance 15-098: The emphasis of the evaluation was on the design, marking, and performance of the weighing load receiving element. A 13' x 2.5', 40 000 x 20 lb, Model LS-WIM Axle Scale weighing element was tested. The scale was interfaced with a Intercomp Company Model PT20 (NTEP CC 12-108) indicating element. The scale was tested using 40 000 lb of known test weights to perform increasing/decreasing, and shift tests. The scale was sealed and used until a minimum time and use requirements had been met, then tested again. The increasing/decreasing load was repeated using 40 000 lb of known test weights.

Evaluated By: D. Flocken (NTEP) 15-098

Type Evaluation Criteria Used: *NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2018 Edition. *NCWM Publication 14 Weighing Devices*, 2018 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 15-098, 15-098A1

Examples of Device:



Model: LS-WIM Axle Scale



Access Cover removed showing junction box and data badge